

Titan[®] Penile Implant Devices

Physician Dictation Reference Sheet

Place any applicable stickers here

Measurement Information		
	Right	Left
Proximal		
Distal		
Total		

Date of Implant		
Approach (circle one)	Infrapubic	Penoscrotal
Model Number		
Serial Number		
Cylinder Size		
Real Tip Extenders	Right: _____	Left: _____
Reservoir Size (circle one)	75 mL	125 mL
Reservoir Fill Volume	_____ cc/mL	
Reservoir Location		

NOTES:

Cylinder Size Titan [®] and Titan [®] Touch	Recommended Fill Volumes	Tubing Length Scrotal (cm/in)	Tubing Length Infrapubic (cm/in)
Zero Degree (0°)			
14 cm Zero Degree (0°)	36 – 41	8.5/3.35	16/6.30
16 cm Zero Degree (0°)	44 – 49	8.5/3.35	16/6.30
18 cm Zero Degree (0°)	54 – 59	10.5/4.13	16/6.30
20 cm Zero Degree (0°)	68 – 73	10.5/4.13	16/6.30
22 cm Zero Degree (0°)	91 – 96	10.5/4.13	16/6.30
Cylinders Only			
24 cm Zero Degree (0°)	107 – 112	Tubing cut to length	
26 cm Zero Degree (0°)	114 – 119	Tubing cut to length	
28 cm Zero Degree (0°)	120 – 125	Tubing cut to length	
Narrow Base			
11 cm Zero Degree (0°)	18 – 23	8.5/3.35	16/6.30
14 cm Zero Degree (0°)	27 – 32	8.5/3.35	16/6.30
16 cm Zero Degree (0°)	32 – 37	8.5/3.35	16/6.30
18 cm Zero Degree (0°)	48 – 53	10.5/4.13	16/6.30

Indications

The Titan, Titan OTR and Titan Touch Inflatable Penile Prosthesis is indicated for male patients suffering from erectile dysfunction (impotence) who are candidates for implantation of a penile prosthesis.

Contraindications

The Titan, Titan OTR and Titan Touch Inflatable Penile Prosthesis is contraindicated in patients with an active infection present anywhere in the body, especially urinary tract or genital infection; with a documented sensitivity to silicone; with unresolved problems affecting urination, such as an elevated residual urine volume secondary to bladder outlet obstruction or neurogenic bladder; or, unwilling to undergo any further surgery for device revision.

Warnings

Implantation of the device may make latent natural erections, as well as other interventional treatment options, impossible.
Men with diabetes or spinal cord injuries, as well as immunocompromised patients, may have an increased risk of infection associated with a prosthesis.
Failure to evaluate and promptly treat erosion may result in a substantial worsening of the condition, leading to infection and loss of tissue.
Implantation of a penile prosthesis may result in penile shortening, curvature or scarring.

Precautions

Surgeons implanting penile prostheses should be familiar with the currently available techniques for measuring the patient, determining implant size, and performing the surgery.
Removal of an implanted prosthesis without timely reimplantation of a new prosthesis may complicate subsequent reimplantation or may make it impossible.

Potential Complications

Potential complications include scrotal swelling, auto-inflation, discomfort, angulation/curvature, edema, device malfunction, chronic pain, difficulty with ejaculation, transient urinary retention.

The information provided is not comprehensive with regard to product risks. For a comprehensive listing of indications, contraindications, warnings, precautions, and adverse events refer to the product's Instructions for Use. Alternatively, you may contact a Coloplast representative at 1-800-258-3476 and/or visit the company Website at www.coloplast.us.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician.